



## RESEARCH

## Open Access

# Impact of mini-implant length on stability at the initial healing period: a controlled clinical study

Manuel Nienkemper<sup>1\*</sup>, Benedict Wilmes<sup>1</sup>, Alexander Pauls<sup>2</sup> and Dieter Drescher<sup>1</sup>

## Abstract

**Introduction:** Aim of the study was to assess the impact of the length of mini-implants inserted in the midpalatal region on the stability at the initial healing period.

**Methods:** A sample of 20 consecutively treated patients ( $15.6 \pm 7.2$  years) was examined. A long mini-implant with a length of 11 mm and a diameter of 2 mm was inserted into the anterior palate of each patient. Resonance frequency analysis (RFA) was performed after insertion (T0), two weeks (T1), four weeks (T2), and six weeks (T3). Insertion depth (ID) and the maximum insertion torque (IT) were measured. RFA, ID and IT data were tested for correlations. RFA values were tested for statistical differences between the different times. Data was compared to a matched control group of patients who received short mini-implants with a length of 9 mm and a diameter of 2 mm.

**Results:** Mean ID was  $9.5 \pm 0.6$  mm and mean IT was  $17.9 \pm 3.8$  Ncm. A correlation was found between RFA and ID ( $r = 0.59$ ,  $P < .01$ ). From T0 to T1 the stability ( $33.4 \pm 3.5$  ISQ) decreased highly significantly by  $5.3 \pm 3.5$  ISQ values ( $P < .001$ ) and significantly from T1 and T2 ( $P < .05$ ) by  $3.5 \pm 3.7$  ISQ values. From T2 on RFA nearly remained unchanged ( $-1.7 \pm 3.9$  ISQ;  $P > .05$ ). At T1 stability was significantly lower than the control group. From T2 on there were no significant differences between the groups.

**Conclusions:** Long mini-implants provide high stability when inserted in the midpalatal region. After initial decrease RFA values remained stable from four weeks on and did not differ from the control group.

**Trial registration:** ID: 2013081293 (Clinical study register, University of Düsseldorf, Germany).

**Keywords:** Skeletal anchorage, Mini-implants, Stability, Length, Healing, Resonance frequency analysis

## Introduction

Anchorage control is a key factor for successful treatment in orthodontics. Skeletal anchorage can be very helpful especially in critical anchorage situations. Dental implants [1], palatal implants [2] or mini-plates [3] were used for this purpose. Mini-implants, introduced by Kanomi, have become increasingly widespread in the recent past because of their low invasiveness during insertion and removal, their versatility and low costs [4]. However, relatively high failure rates ranging from 10% up to 30% remain the major problem using this type of temporary anchorage devices [5]. For that reason researchers have been working on various designs in order to optimize mini-implant stability:

There are some studies assessing the influence of different materials [6] or surface treatment [7] to enhance cellular reaction and micro retention. Other studies focus on macroscopic design elements such as thread and taper design [8]. Different lengths and diameters as well as conical or cylindric shapes were tested for their influence on primary stability [9,10]. Most of these investigations are *in-vitro* studies. In the clinical studies insertion torque is mostly used as a measure for primary stability. The validity of this parameter to assess implant stability and to predict success, however, is still controversial [11]. In addition it only allows assessing initial stability but not secondary stability which is even of more relevance for clinical success.

Up to now, all the studies investigating the influence of different mini-implant designs are confined to primary stability measurement. Without doubt, primary stability is a basic prerequisite for clinical

\* Correspondence: [Manuel.Nienkemper@uni-duesseldorf.de](mailto:Manuel.Nienkemper@uni-duesseldorf.de)

<sup>1</sup>Department of Orthodontics, Heinrich-Heine-University, Moorenstr. 5, 40225 Düsseldorf, Germany

Full list of author information is available at the end of the article

**Table 1 Comparison regarding gender and age between the groups**

	2 × 11 mm		2 × 9 mm			
Gender:	10 male	10 female	11 male	8 female	n.s.	Chi-square-test
Age (years):	15.62	7.19	15.55	7.34	n.s.	Mann-Whitney-U-Test
* p < .05    **p < .001    ***p < .0001						

success [12]. But due to the remodelling process that occurs at the bone to implant interface during the healing period, stability is significantly affected [13]. Hence, it would be highly interesting to investigate the influence of different design features on primary *and* secondary stability of orthodontic mini-implants [14,15]. In dental implantology resonance frequency analysis (RFA) is established for clinical stability measurement subsequent to insertion and at any time during further treatment [16,17]. In the recent past RFA has proven to be suitable also for assessment of mini-implant stability [18].

The aim of this clinical study was to determine the stability of long mini-implants with a length of 11 mm at the initial healing period of six weeks and to compare it with the results of shorter mini-implants with a length of 9 mm derived from a previous pilot study [19].

Materials and methods

Subjects

Only patients whose treatment plan comprised the insertion of palatal mini-implants were taken into consideration. Absence of systemic diseases affecting bone metabolism or wound healing and good oral hygiene were

further inclusion criteria. Patients who missed examination appointments or showed signs of peri-implant inflammation were excluded. Participation was non-mandatory, each patient signed a patient consent form.

Power analysis

For sample size calculation data of a clinical pilot study was used. This study investigated the stability at the initial healing period of palatally inserted mini-implants with a length of 9 mm, based on the same observation protocol. Significant changes in stability were found between the second and the fourth week. Consequently, ISQ changes and standard deviations of this observation period were used for calculation.

The changes in ISQ values of 7.89 with a standard deviation of 5.92, a chosen alpha level of 0.001 and a power of 0.95 resulted in a required sample size for the treated and the control group of n = 19. G\*Power 3.1.5 (University of Kiel, Germany) software was chosen for this calculation. During the pilot study 4 patients had to be excluded either due to signs of peri-implant inflammation or missed examination appointments. As a consequence 23 consecutively treated patients were examined for this study. No further selection was performed.

Having 3 drop outs the test group finally comprised 20 patients, 10 males and 10 females of white ancestry with a mean age of 15.6 ± 7.2 years (Table 1).

The prospective study was approved by the ethical committee of the University Clinics of Düsseldorf, Germany. It was performed according to the Declaration of Helsinki guidelines on experimentation involving human subjects.

Insertion protocol

Benefit mini-implants (length: 11 mm, outer diameter: 2 mm, inner diameter: 1.35 mm, thread pitch: 0.75 mm)



**Figure 1** 2 × 11 mm mini-implant inserted in the median part of the anterior palate between the second and third palatine rugae.

(PSM medical solutions; Tuttlingen, Germany) were inserted in the midpalatal suture between the second and third palatine rugae (Figure 1). After predrilling using a 1.3 mm drill to a depth of 3 mm insertion was performed rectangular to the palatal curvature until the mini-implant's head touched the soft tissue. A surgical machine (ElcoMed SA 200C, W&H, Bürmoos, Austria) was used for predrilling and insertion.

### Observation protocol

Prior to insertion the soft tissue thickness (STT) was determined using a dental probe with a rubber stop. Insertion depth (ID) was calculated subtracting these values from the mini-implants' thread length (11 mm). During insertion the maximum insertion torques value (IT) were recorded by the surgical machine. After insertion the RFA was performed three times along the midpalatal suture and three times perpendicular to it (T0). RFA measurements were repeated after 2 (T1), 4 (T2) and 6 (T3) weeks. Mean ISQ values were calculated from the repeated measurements in both directions and as overall values.

Peri-implant soft tissue was each time observed for signs of inflammation.

### Matching of the control group

The data was additionally compared to those of the already mentioned pilot-study. In that study the same observation protocol was used to evaluate the changes in stability of 2 × 9 mm Benefit mini-implants. The control group consists of 19 patients, 11 males and 8 females of white ancestry with a mean age of 15.5 ± 7.3 years. The control group matched the treatment group as to sex distribution and age (Table 1).

The only differences between the two groups were the length of the mini-implants and the spot for insertion, which is a little bit more anterior in this study.

### Statistical analysis

Normal distribution (Shapiro-Wilk test) as well as equality of variances (Levene-test) were found for all ISQ and ID values. Only IT values did not show normal distribution. Hence ANOVA and Duncan post hoc test as parametric tests were applied to assess differences between

**Table 3 Correlations between RFA, insertion torque (IT) and soft tissue thickness (ST) (2 × 11 mm)**

	r	p	r <sup>2</sup>
ID - IT	0.31	0.18	0.1
RFA-IT	0.26	0.27	0.07
RFA-ID	0.59	<0.01	0.35

the ISQ values at T0, T1, T2 and T3. For comparison of ISQ values parallel and perpendicular to the midpalatal suture the paired *t*-test was used. Pearson correlation and linear regression analysis were chosen to analyze relations between initial ISQ, IT, and ID values.

For inter-group comparison unpaired *t*-test was used for differences in ISQ values at each observation appointment and in ID values. For comparison of IT values Mann-Whitney-*U*-test was chosen since the respective data did not show normal distribution for the 2 × 11 mm mini-implants.

Statistics were performed with the statistical software SPSS 21.0 (IBM, Chicago, Ill). Statistical significances were tested at *P* < .05 (\*), *P* < .001 (\*\*) and *P* < .0001 (\*\*\*) levels.

### Results

Soft tissue thicknesses at the insertion site were found to be nearly identical in both groups (STT: 1.55 ± 0.60 mm vs. 1.50 ± 0.55 mm). Given the higher length of the implants used in the present study the insert depth was nearly 2 mm higher than in the pilot study (ID: 9.45 ± 0.60 vs. 7.50 ± 0.55) (Table 2).

Mean IT was 17.85 Ncm and mean initial ISQ was 33.35 ± 3.53. In relation to the control group IT was slightly higher whereas RFA showed lower values. Differences did not reach the level of significance.

Within the test group a significant correlation could be found between ID and initial ISQ values with a correlation coefficient of *r* = 0.59 (Table 3). Between ID and IT and between RFA and IT there was no correlation.

The stability shown by RFA measurements was subject to changes over the first six weeks. The initial ISQ of 33.35 ± 3.53 decreased significantly during the first two weeks by 5.25 ± 3.53 (*p* < .001) and during week three and four by 3.45 ± 3.66 (*p* < .05) (Tables 4 and 5); Figure 2. From week

**Table 2 Initial values for RFA, insertion torque (IT) and soft tissue thickness (ST)**

	2 × 11 mm		2 × 9 mm		p	
	mean	SD	mean	SD		
STT (mm)	1.55	0.6	1.5	0.55	n.s.	<i>t</i> -test
ID (mm)	9.45	0.60	7.5	0.55	***	<i>t</i> -test
IT (Nmm)	17.85	3.75	16.81	3.54	n.s.	<i>U</i> -test
RFA (ISQ)	33.35	3.53	36.14	6.08	n.s.	<i>t</i> -test
* <i>p</i> < .05	** <i>p</i> < .001	*** <i>p</i> < .0001				

**Table 4 RFA values (ISQ) at each measuring point (2 × 11 mm)**

	T0		T1		T2		T3	
	mean	SD	mean	SD	mean	SD	mean	SD
Overall	<b>33.35</b>	3.53	<b>28.10</b>	3.99	<b>24.63</b>	4.46	<b>22.90</b>	6.00
length	33.55	3.92	28.35	4.43	24.62	4.39	23.30	5.90
square	33.15	3.27	27.85	3.91	24.65	4.82	22.50	6.30
	ns		ns		ns		ns	
* <i>p</i> < .05	** <i>p</i> < .001	*** <i>p</i> < .0001						

**Table 5 Changes in mini-implant stability over time measured by RFA (ISQ)**

	T1-T0		T2-T1		T3-T2		T3-T0	
	mean	SD	mean	SD	mean	SD	mean	SD
Test (2x11mm)	-5.25	3.53	-3.45	3.66	-1.73	3.88	-10.45	5.22
	**		*		ns		***	
Control (2x9mm)	-4.03	6.08	-7.89	5.92	-1.72	3.53	-13.63	9.49
	ns		**		ns		***	
	* p < .05	**p < .001	***p < .0001					

four on the stability remained nearly unchanged. Overall the stability decreased significantly by  $10.45 \pm 5.22$  ISQ to a mean level of  $22.9 \pm 6.0$ . There were no statistical differences regarding measurement direction.

The comparison with the control group showed significant lower stability at T1 after comparable initial ISQ values due to the significant decrease of stability from T0 to T1 in the test group ( $p < .001$ ) (Table 6, Figure 3). From T1 to T2 there is a slight decrease ( $p < .05$ ) whereas ISQ values in the control group decrease by  $7.89 \pm 5.92$  ( $p < .001$ ). As a result the stability at T2 is nearly the same in both groups. From T2 to T3 stability remains nearly unchanged in both groups.

## Discussion


For this prospective clinical study a sufficient number of patients could be included after drop outs, as calculated by power analysis on the basis of a clinical investigation using a nearly identical procedure [19]. The patients of that pilot study were found to be suitable as control group to identify the effects of implant length on the primary and secondary stability, since they match the test group as to age and

gender. A test for normal distribution for small sample sizes was used to select appropriate statistics. The measurement methods used for stability assessment have proven to be accurate enough for this purpose [18,20].

The results indicated a high primary stability as shown by the RFA and IT values at T0. Interestingly, the stability was not higher than those of the 9 mm long mini-implants of the control group. In many studies stability is determined by measurement of maximum insertion torque [14,15].

A closer look into the literature addressing the impact of implant length reveals that different measurement methods and implant loading modes may characterize different aspects of implant stability:

Pithon et al. reported higher maximum insertion torque values, when longer mini-implants were used [21]. The current results show the same tendency, but yet not significantly. But in that *in vitro* study higher implant length did not lead to higher mechanical resistance when loaded by a lateral force. Chatzigianni et al. found that the mini-implant's length only affected lateral displacement at a particular force value



**Figure 2 Stability of 2 × 11 mm mini-implants over six weeks by means of RFA (ISQ).**

**Table 6 Comparison of RFA values (ISQ) at each measuring point**

	T0		T1		T2		T3	
	mean	SD	mean	SD	mean	SD	mean	SD
Test (2x11 mm)	<b>33.35</b>	3.53	<b>28.1</b>	3.99	<b>24.63</b>	4.46	<b>22.9</b>	6
Control (2x9 mm)	<b>36.14</b>	6.08	<b>32.11</b>	5.57	<b>24.23</b>	7.19	<b>22.51</b>	6.69
Difference	<b>-2.79</b>		<b>-4.01</b>		<b>0.4</b>		<b>0.39</b>	
p	0.086		0.013		0.833		0.848	
	n.s.		*		n.s.		n.s.	
	* p < .05	**p < .001	***p < .0001					


[22]. In our study the ISQ values of the longer mini-implants were lower than those of the control group but without reaching the level of significance. Pan et al. reported of lower ISQ values for longer mini-implants and stated that insertion depth related to the absolute implant length is of higher importance for stability and mechanical load capacity. The current investigation also found a correlation between RFA and ID whereas no correlation was found between RFA or ID to IT. The fact that orthodontic mini-implants are primarily loaded by lateral forces suggests that RFA might be more suitable to detect this kind of stability than IT.

The direction of RFA measurement after insertion at the midpalatal region showed no significant differences as it was already presumed by the results of the pilot-study.

The position of the mini-implants might have a minor impact on stability. The longer mini-implants were inserted slightly more anteriorly than in the control group (directly distal from the third palatine rugae) due to greater bone height in this area [23]. They consequently came closer to

the incisive canal. Song et al. investigated the microanatomy of the incisive canal and found that there is a high variation in shape, diameter and course. Even the number of canals ranged from one to four. As a result, no guidelines can be given to the clinician how to reliably prevent touching or penetrating the canal. But the higher length, the position and axis of the mini-implants superimposed with the anticipated path of the incisive canal give reason to presume that the canal was hit in some cases. Somehow, this might affect mini-implant stability. The authors also found that there were numerous arteries and veins as well as bundles of small nerves in relatively large canals. Since there is plenty of space within the canal, the small nerve fibres may easily slip away when a mini-implant penetrates. Hence, sequelae such as numbness of the anterior palatal mucosa due to damage of the nasopalatine nerve are rather unlikely and were not observed, yet. Also in the present and the pilot study none of the patients complained about numbness at the respective soft tissues.

The current investigation suggests that longer, i.e. 11 mm mini-implants in the anterior palate that have



**Figure 3 Comparison of the stability of 2 × 11 mm and 2 × 9 mm mini-implants over six weeks by means of RFA (ISQ).**



to be inserted slightly more anteriorly due to bone height do not provide higher stability than 9 mm mini-implants, neither immediately after insertion nor after the initial healing period of six weeks. After two weeks stability was even lower. This is due to a significant decrease ( $-5.25 \pm 3.53$  ISQ;  $p < .001$ ) from T0 to T1 whereas the shorter implants of the control group showed no significant stability changes. During T1 and T2 when the control group had the major loss of stability ( $-7.89 \pm 5.92$  ISQ;  $p < .001$ ) the decrease of the  $2 \times 11$  mm mini-implants flattened, but was still significant ( $3.45 \pm 3.66$ ;  $p < .05$ ). From week four on the stability remained constant in both groups showing no significant differences.

Overall the development of stability was similar in both groups. Stability decreased significantly for the first four weeks in accordance to dental implants that showed lowest stability after three [24] or four [25] weeks. Stability of surface treated implants already had begun to increase after six weeks [26] whereas mini-implant's stability of both groups remained unchanged. For orthodontic purposes this level of stability seems to be suitable. It also should be kept in mind that excessively high stability levels may be detrimental, since temporary anchorage devices of small diameter might break at the time of removal [27].

By now, there is no study evaluating the influence of mini-implant length on secondary stability. Sim et al. investigated the development of stability of dental implants 8 mm and 10 mm in length [28]. Over the first two weeks initial differences in stability were found that did not reach the level of significance due to high standard deviations. After that stability developed nearly identically independent from implant length.

Further clinical investigations regarding the factors that may affect primary and secondary stability such as different mini-implant diameters, surface treatment or different insertion sites should be performed to optimize the clinical protocols and success rates of skeletal anchorage.

## Conclusions

Long 11 mm mini-implants provide a high level of stability when inserted at the midpalatal region.

After initial loss the stability did not change significantly from week four on. After the first two weeks the longer mini-implants inserted in a more anterior position even showed a lower stability than 9 mm long mini-implants. After four weeks the stability was comparable. Regarding stability there seems to be no advantage in using longer mini-implants at the median part of the palate. Preferring short implants of 9 mm length inserted slightly posteriorly may also help to

prevent hitting the incisive canal even though this might not be of clinical relevance.

## Abbreviations

RFA: Resonance frequency analysis; ISQ: Implant stability quotient; IT: Maximum insertion torque value; ID: Insertion depth; STT: Soft tissue thickness.

## Competing interests

Dr. Wilmes is the inventor of the Benefit System.

## Authors' contributions

MN performed the measurements, statistical analysis and drafted the manuscript. AP coordinated patients' appointments and took part in measuring. DD and BW took part in writing the paper. All authors read and approved the final manuscript.

## Author details

<sup>1</sup>Department of Orthodontics, Heinrich-Heine-University, Moorenstr. 5, 40225 Düsseldorf, Germany. <sup>2</sup>Orthodontist, private practice, Gersthofer Str. 2, 65929 Frankfurt am Main, Germany.

Received: 20 August 2013 Accepted: 9 October 2013

Published: 20 October 2013

## References

- Kokich VG: Managing complex orthodontic problems: the use of implants for anchorage. *Semin Orthod* 1996, **2**:153–160.
- Triaca A, Antonini M, Wintermantel E: Ein neues titan-Flachschrauben-Implantat zur orthodontischen Verankerung am anterioren Gaumen. *Informationen aus Orthodontie und Kieferorthopädie* 1992, **24**:251–257.
- Umemori M, Sugawara J, Mitani H, Nagasaka H, Kawamura H: Skeletal anchorage system for open-bite correction. *Am J Orthod Dentofacial Orthop* 1999, **115**:166–174.
- Kanomi R: Mini-implant for orthodontic anchorage. *J Clin Orthod* 1997, **31**:763–767.
- Reynders R, Ronchi L, Bipat S: Mini-implants in orthodontics: a systematic review of the literature. *Am J Orthod Dentofacial Orthop* 2009, **135**:564 e561–519. discussion 564–565.
- Pan CY, Chou ST, Tseng YC, Yang YH, Wu CY, Lan TH, Liu PH, Chang HP: Influence of different implant materials on the primary stability of orthodontic mini-implants. *Kaohsiung J Med Sci* 2012, **28**:673–678.
- Kim SH, Cho JH, Chung KR, Kook YA, Nelson G: Removal torque values of surface-treated mini-implants after loading. *Am J Orthod Dentofacial Orthop* 2008, **134**:36–43.
- Chang JZ, Chen YJ, Tung YY, Chiang YY, Lai EH, Chen WP, Lin CP: Effects of thread depth, taper shape, and taper length on the mechanical properties of mini-implants. *Am J Orthod Dentofacial Orthop* 2012, **141**:279–288.
- Holm L, Cunningham SJ, Petrie A, Cousley RR: An *in vitro* study of factors affecting the primary stability of orthodontic mini-implants. *Angle Orthod* 2012, **82**:1022–1028.
- Kim JW, Baek SH, Kim TW, Chang YI: Comparison of stability between cylindrical and conical type mini-implants. Mechanical and histological properties. *Angle Orthod* 2008, **78**:692–698.
- Meursing Reynders RA, Ronchi L, Ladu L, van Etten-Jamaludin F, Bipat S: Insertion torque and success of orthodontic mini-implants: a systematic review. *Am J Orthod Dentofacial Orthop* 2012, **142**:596–614. e595.
- Meredith N: Assessment of implant stability as a prognostic determinant. *Int J Prosthodont* 1998, **11**:491–501.
- Gedrange T, Proff P, Bayerlein T, Landsberger P, Dietze S, Fanghanel J: Histological and fluorescence microscopic examination of the bone/implant interface in orthodontic miniscrews (Mondeal). *Folia Morphol (Warsz)* 2006, **65**:70–71.
- Javed F, Almas K, Crespi R, Romanos GE: Implant surface morphology and primary stability: is there a connection? *Implant Dent* 2011, **20**:40–46.
- Wilmes B, Rademacher C, Olthoff G, Drescher D: Parameters affecting primary stability of orthodontic mini-implants. *J Orofac Orthop* 2006, **67**:162–174.
- Meredith N, Alleyne D, Cawley P: Quantitative determination of the stability of the implant-tissue interface using resonance frequency analysis. *Clin Oral Implants Res* 1996, **7**:261–267.

17. Quesada-García MP, Prados-Sánchez E, Olmedo-Gaya MV, Muñoz-Soto E, González-Rodríguez MP, Vallecillo-Capilla M: **Measurement of dental implant stability by resonance frequency analysis: a review of the literature.** *Med Oral Patol Oral Cir Bucal* 2009, **14**:e538–e546.
18. Nienkemper M, Wilmes B, Panayotidis A, Pauls A, Golubovic V, Schwarz F, Drescher D: **Measurement of mini-implant stability using resonance frequency analysis.** *Angle Orthod* 2012, **83**:230–238.
19. Nienkemper M, Wilmes B, Pauls A, Drescher D: **Mini-implant stability at the initial healing period.** *Angle Orthod* 2013. Epub ahead of print.
20. Pauls A, Nienkemper M, Drescher D: **Accuracy of torque-limiting devices used for mini-implant placement—an *in vitro* study.** *J Orofacial Orthopedics* 2013. In press.
21. Pithon MM, Figueiredo DS, Oliveira DD: **Mechanical evaluation of orthodontic mini-implants of different lengths.** *J Oral Maxillofac Surg* 2013, **71**:479–486.
22. Chatzigianni A, Keilig L, Reimann S, Eliades T, Bourauel C: **Effect of mini-implant length and diameter on primary stability under loading with two force levels.** *Eur J Orthod* 2011, **33**:381–387.
23. Ludwig B, Glasl B, Bowman SJ, Wilmes B, Kinzinger GS, Lisson JA: **Anatomical guidelines for miniscrew insertion: palatal sites.** *J Clin Orthod* 2011, **45**:433–441. Quiz 467.
24. Han J, Lulic M, Lang NP: **Factors influencing resonance frequency analysis assessed by Osstell mentor during implant tissue integration: II. Implant surface modifications and implant diameter.** *Clin Oral Implants Res* 2010, **21**:605–611.
25. Huwiler MA, Pjetursson BE, Bosshardt DD, Salvi GE, Lang NP: **Resonance frequency analysis in relation to jawbone characteristics and during early healing of implant installation.** *Clin Oral Implants Res* 2007, **18**:275–280.
26. Sul YT, Jonsson J, Yoon GS, Johansson C: **Resonance frequency measurements *in vivo* and related surface properties of magnesium-incorporated, micropatterned and magnesium-incorporated TiUnite, Osseotite, SLA and TiOblast implants.** *Clin Oral Implants Res* 2009, **20**:1146–1155.
27. Barros SE, Janson G, Chiqueto K, Garib DG, Janson M: **Effect of mini-implant diameter on fracture risk and self-drilling efficacy.** *Am J Orthod Dentofacial Orthop* 2011, **140**:e181–e192.
28. Sim CP, Lang NP: **Factors influencing resonance frequency analysis assessed by Osstell mentor during implant tissue integration: I. Instrument positioning, bone structure, implant length.** *Clin Oral Implants Res* 2009, **21**:598–604.

doi:10.1186/1746-160X-9-30

**Cite this article as:** Nienkemper et al.: Impact of mini-implant length on stability at the initial healing period: a controlled clinical study. *Head & Face Medicine* 2013 **9**:30.

**Submit your next manuscript to BioMed Central and take full advantage of:**

- Convenient online submission
- Thorough peer review
- No space constraints or color figure charges
- Immediate publication on acceptance
- Inclusion in PubMed, CAS, Scopus and Google Scholar
- Research which is freely available for redistribution

Submit your manuscript at  
www.biomedcentral.com/submit

